



PINNACLE — ANALYTICS —

Potency Results

Sample Name: Master OG
Client: Earthy Now
Client Batch ID:

Pinnacle-Analytics.com
3549 Lear Way, Suite 101
Medford OR 97504
P:(541)300-8217

Sample ID: rC-H-390-E2432

Matrix: Flower

Prep Analyst: Jeff A.

Analysis Method: 0668534+1 H4 5-24-2024 #1.lcm

Sampling Method: N/A

Reference Method: JCB 2009: HPLC/DAD

Analysis Batch: 11-6-2024 H4 302, 390 Flower

Date Sampled: 11/5/2024

Date Reported: 11/8/2024

Client License: AG-R1047246IHG

14010 Hwy 62

Eagle Point OR 97524


For R&D Purposes Only

Total THC (THCA*0.877+d9-THC) **0.579%**
Total CBD (CBDA*0.877+CBD) **14.9%**
Moisture Content **12.1%**

Cannabinoid	% Weight	mg/g
CBDVA	0.129	1.29
CBDV	<LOQ	<LOQ
CBDA*	16.5	165.0
CBGA	0.479	4.79
CBG	<LOQ	<LOQ
CBD*	0.468	4.68
THCV	<LOQ	<LOQ
CBN	<LOQ	<LOQ
d9-THC*	<LOQ	<LOQ
d8-THC*	<LOQ	<LOQ
CBC	<LOQ	<LOQ
THCA*	0.66	6.6
Total Cannabinoids	18.24	182.0

*ORELAP Accredited Analyte

Limit Of Quantitation: 0.1%, analyte not measured



- CBDVA
- CBD*
- CBDA*
- THCA*
- CBGA



These test results may not be altered or reproduced except in full without the permission of Pinnacle Analytics. These results were generated following the Oregon Administrative Rules and in accordance with the NELAP Institute under ORELAP License #4152 Report generated by Routine_Potency_Rev13_8-1-2023

Kris Ford, PhD
Lab Director



PINNACLE ANALYTICS

Quality Control Results

Analyst: Jeff A.

Analysis Batch: 11-6-2024 H4 302, 390 Flower

Pinnacle-Analytics.com
3549 Lear Way, Suite 101
Medford OR 97504
P:(541)300-8217

	Duplicate RPD		LCS % Recovery		Method Blank	
	H-0-E2431-b	Limit	C-FL-110624	Limits	C-FB-110624	Limit
CBDA	3.11%	10%	98.6%	90-110%	<LOQ/2	LOQ/2
CBD	2.67%	30%	97.5%	90-110%	<LOQ/2	LOQ/2
d9-THC	<LOQ%	30%	100.0%	90-110%	<LOQ/2	LOQ/2
d8-THC	<LOQ%	30%	96.0%	90-110%	<LOQ/2	LOQ/2
THCA	3.81%	10%	99.5%	90-110%	<LOQ/2	LOQ/2

RPD: Relative Percent Difference between unknown sample and its duplicate

LCS: Laboratory Control Sample with known concentration

Case Comments: There were no divergences from ordinary Quality Control procedures or SOPs.



These test results may not be altered or reproduced except in full without the permission of Pinnacle Analytics. These results were generated following the Oregon Administrative Rules and in accordance with the NELAP Institute under ORELAP License #4152
Report generated by Routine_Potency_Rev13_8-1-2023

Kris Ford, PhD
Lab Director